

TOPLEVEL PARAMETERS

Method Information For: D:\MSDCHEM\1\METHODS\MAN27.M

Method Sections To Run:

- (X) Save Copy of Method With Data
- () Pre-Run Cmd/Macro =
- (X) Data Acquisition
- (X) Data Analysis
- () Post-Run Cmd/Macro =

Method Comments:

Quantification du rapport Testosterone /Epitestosterone MSD20 injection
en split

END OF TOPLEVEL PARAMETERS

INSTRUMENT CONTROL PARAMETERS

Sample Inlet: GC
Injection Source: GC ALS
Mass Spectrometer: Enabled

=====
== 6890 GC METHOD
=====

OVEN

Initial temp: 160 'C (On)
Initial time: 0.00 min

Maximum temp: 325 'C
Equilibration time: 0.50 min

Ramps:

#	Rate	Final temp	Final time
1	4.00	255	0.00
2	30.00	300	2.75
3	0.0(Off)		

Post temp: 0 'C
Post time: 0.00 min
Run time: 28.00 min

FRONT INLET (UNKNOWN)

Mode: Split
Initial temp: 280 'C (On)
Pressure: 170.0 kPa (On)
Split ratio: 12.1:1
Split flow: 11.3 mL/min
Total flow: 14.8 mL/min
Gas saver: Off
Gas type: Helium

BACK INLET ()

COLUMN 1

COLUMN 2

Method: MAN27.M

Thu Aug 03 14:52:27 2006

Page:1

USADA 0265

3